

A Government of India Enterprise
RADIATION DETECTORS & INSTRUMENTATION DIVISION-PURCHASE
ECIL Post, Hyderabad - 500 062. (A. P)

Your offer should reach us on or before the Due Date.

e-mail:

If you have 'NO BID' kindly send your 'REGRET' Letter for updating our vendor Directory

FOR ELECTRONICS CORPORATION OF INDIA LIMITED

Yours faithfully,

CORPORATION OF INDIA
09-11-2013

INSTRUCTION AND COMMERCIAL CONDITIONS:

ENQUIRY: 340305

1. The quote is to be submitted on or before the due date mentioned.
2. The BID is of TWO PART BID Mechanism.
Viz: Technical Bid (Part A) and Price Bid (Part-B)


TECHNICAL BID SHALL CONTAIN:

- a) Technical Data.
- b) Unpriced Bid.
- c) Compliance to our commercial conditions.
- d) Compliance to our Technical Specifications.

⑥ details of similar orders executed.

THE PRICE BID SHALL CONTAIN:

- a) PRICE
3. We require validity of 30 days from the date of opening of Price Bid.
 4. We require the Delivery of items within 6-8 weeks of receipt of purchase order by you.
 5. Please confirm about delivery schedule in your offer. The Standard Liquidated Damages clause is applicable for Delay in Deliveries. The Clause Read as " In the Event of any delay in deliveries beyond agreed schedule, Liquidated Damages will Be recovered @ 1/2 % per week subject to a maximum of 5% on undelivered portion".
 6. Payment terms: 100% of total order value will be made within 60 days of receipt of acceptance.
 7. The standard risk purchase clause is applicable to this requirement. The clause read as "In the event of failure to comply with the purchase order terms & conditions, ECIL has right to cancel the order and proceed with an alternate source. In such case, you shall have to bear the extra cost incurred by us which will be recovered from your available credit". Please convey your acceptance in your quote.
 8. We require warranty of 12 months from the date of supply. Please explain modus operandi for warranty support. The replacements for defective items shall be done within two weeks of intimation from ECIL.
 9. You need to submit Security Deposit as well as Performance Guarantee for a value of 10% to cover execution and warranty period in case of order on you.
 10. Certificate of compliance in triplicate shall be submitted along with dispatches stating that the material supplied are in accordance with the specifications relevant to this order.
 11. Documentation including maintenance & operating manual shall be supplied along with Material. (1 no. for each instrument, both in Hard copy as well as soft copy). Please explain in this regard.
 12. Installation & Commissioning shall be done at ECIL works at free of cost.
 13. Replacement to rejections if any shall be supplied with all costs to your account till matériel reaches ECIL.
 14. Please comment on each of above clauses and on technical specification enclosed, in your Offer without fail.


09-11-2013

वाई.वी.वी.एस.बी. बोस
Y.V.V.S.B. BOSE
उप महा प्रबन्धक/Dy. General Manager
आर आर टी डिवीजन सी आई एल
R2/PURCHASE/ECIL
हैदराबाद/HYDERABAD-500 052.

TURBO PACK CONSIST OF FOLLOWING ITEMS

S.NO	MATERIAL DESCRIPTION	QTY.
1.	Corrosive version Turbo Molecular Pump with Display Control unit, air cooling kit, Connection cables, DN-160 to DN-16 flange fittings set flexible SS304L pipes and necessary accessories with standard test certificates. <u>Specifications:-</u> ➤ Pumping speed : 425-450 lps ➤ Ultimate Pressure : $< 1 \times 10^{-8}$ mbar.	6 sets.
2.	a) Corrosive Version double stage Rotary vacuum pumps with suitable oil with standard test certificates. <u>Specifications:-</u> ➤ Pumping speed : $20 \text{ m}^3/\text{hr}$ ➤ Ultimate Pressure : $< 6 \times 10^{-3}$ mbar	8 Nos.
	b) Corrosive Version Dry vacuum pumps <u>Specifications:-</u> ➤ Pumping speed : $\approx 20 \text{ m}^3/\text{hr}$ ➤ Ultimate Pressure : $< 2 \times 10^{-2}$ mbar.	2 Nos.
3.	Vacuum Gauges system consist of Active pirani/cold cathode transmitter with works calibration certificate. <u>Specifications:-</u> ➤ Measuring Range : Atmospheric pressure to 1×10^{-8} mbar.	
	a) Corrosive version Gauge Heads with sensor cable	10 Sets.
	b) Two channel Vacuum Display and Control Unit with Power cards	5 Sets.
4.	Suitable Trolley for Turbo Pack	5 Sets.
5.	Venting Valves	5 Nos.
6.	Oil Misters	5 Nos.
7.	Electro-magnetic NW-16 Flange valves	10 Nos

Handed over
28/08/13.