

Telefax : +91-40-27146073
Phone : +91-40-27182714, 27182651
E Mail : ybbose@ecil.co.in

Our Enquiry REF: EC/RID/PUR/ : 345877:FE:T: 916

DATE: 2013.09.02 DUE DATE: 2013.09.16

Contact: YVVS.B. BOSE
Dear Sirs,

Please submit your Quotation in a SEALED envelope Superscribing
our Enquiry reference & Due Date for the supply of the following items.

Your offer should reach us on or before the Due Date.



इलेक्ट्रॉनिक्स कारपोरेशन ऑफ इण्डिया लिमिटेड
Electronics Corporation of India Limited

A Government of India Enterprise
RADIATION DETECTORS & INSTRUMENTATION DIVISION-PURCHASE
ECIL Post, Hyderabad - 500 062. (A. P)

EXTRA COPY

Tel No.:
e-mail:

S.No.	DESCRIPTION	UNIT CODE	QUANTITY
1	QPL QUALIFIED PANCAKE GM DETECTOR ALONG WITH ANODE CLIP (PRODUCT SHALL MEET ALL THE SPECIFICATIONS & CLIMATIC,VIBRATION ETC STANDARDS AS SPECIFIED/APPLICABLE FOR QPL).	Sets	1,400.000

Special Notes.

1. PLEASE QUOTE YOUR BEST OFFER IMMEDIATELY. (INDIAN ORIGIN VENDORS SHALL SUBMIT QUOTES IN COVERS WITH PROPER SEALING COVERS RECEIVED WITHOUT PROPER SEALING WILL BE REJECTED).
2. OUR STANDARD PAYMENT TERMS ARE 60 DAYS CREDIT.
3. PLEASE QUOTE FOR FREE DELIVERY AT OUR STORES.
4. CERTIFICATE OF COMPLIANCE FROM MANUFACTURER STATING THAT THE SUPPLIED MATERIAL CONFORMS TO SPECIFICATIONS RELEVANT TO THIS PROCUREMENT SHALL BE SUBMITTED WITH THE

If you have 'NO BID' kindly send your 'REGRET' Letter
for updating our vendor Directory

Yours faithfully,
For ELECTRONICS CORPORATION OF INDIA LIMITED

08-09-2013
Y.V.V.S.B. BOSE
उप महा प्रबंधक/DY. General Manager
आर और डी/आर/ई सी आई एल
RID/PURCHASE/ECIL
हैदराबाद/HYDRABAD-500 062.

Telefax : +91-40-27146073
Phone : +91-40-27182714, 27182651
E Mail : ybbose@ecil.co.in

Our Enquiry REF: EC/RID/PUR/

Enquiry No : 345877

DATE:

DUE DATE :

Dear Sirs,

Please submit your Quotation in a SEALED envelope Superscribing
our Enquiry reference & Due Date for the supply of the following items.

Your offer should reach us on or before the Due Date.

S.No.	DESCRIPTION	UNIT CODE	QUANTITY
	<ul style="list-style-type: none">MATERIAL.5. YOU SHALL ENSURE THAT THE MARKING OF MANUFACTURER'S NAME PART NO. OF ITEM & DATE/BATCH CODE ON THE COMPONENT (DATE/BATCH CODE SHALL BE WITHIN TWO YEAR FROM THE DATE OF ORDER) AND INDICATE THE SAME IN THE INVOICE, PACKING LIST AND COC.6. MATERIAL REQUIREMENT IS URGENT. PLEASE QUOTE YOUR BEST DELIVERY SCHEDULE.7. WARRANTY OF 2 YEARS IS REQUIRED.8. YOU SHALL SUBMIT PERFORMANCE BANK GUARANTEE FOR 10% OF ORDER VALUE AND VALID TILL WARRANTY WHILE CLAIMING PAYMENT.9. ISI PROCEDURE SHALL BE SUBMITTED WITH THE QUOTE.10. TEST REPORTS AND CALIBRATION CERTIFICATE SHALL BE SUBMITTED BEFORE EFFECTING THE DISPATCH IN CASE OF ORDER ON YOU.11. PLEASE COMPLY ALL ABOVE POINTS IN YOUR QUOTE.		

If you have 'NO BID' kindly send your 'REGRET' Letter
for updating our vendor Directory



इलेक्ट्रॉनिक्स कारपोरेशन ऑफ़ इण्डिया लिमिटेड
Electronics Corporation of India Limited

A Government of India Enterprise
RADIATION DETECTORS & INSTRUMENTATION DIVISION-PURCHASE
ECIL, Post, Hyderabad - 500 062. (A. P)

34F D6 003 R0

Yours faithfully,
For ELECTRONICS CORPORATION OF INDIA LIMITED

02.09.2013
वा.वी.बी.एस.बी. बोस
Y.V.V.S.B. BOSE
उप महा प्रबन्धक/Dy. General Manager
आर और ई/अंश/ई टी और एल
RID/PURCHASE/ECIL
हैदराबाद/HYDRABAD-500 062.

QPL QUALIFIED

PANCAKE MICA WINDOW-ALPHA-BETA-GAMMA DETECTOR

GENERAL SPECIFICATIONS

Gas Filling	Ne +Halogen
Cathode Material	446 Stainless Steel
Maximum Length (inch/mm)	3.00 / 76.1
Effective Depth (inch/mm)	0.5 / 12.7
Maximum Diameter (inch/mm)	2.11 / 53.6
Effective Diameter (inch/mm)	1.75 / 44.5
Connector	Grid Cap
Operating Temperature Range °C	-55 to +75

WINDOW SPECIFICATIONS

Areal Density (mg/cm ²)	2.0
Effective Diameter (inch/mm)	1.75 / 44.5
Material	Mica

ELECTRICAL SPECIFICATIONS

Recommended Anode Resistor (meg ohm) (Recommended circuit should be provided along with quotation)	3.3Mohms
Maximum Starting Voltage (volts)	750
Recommended Operating Voltage (volts)	900
Operating Voltage Range (volts)	850-1000
Maximum Plateau Slope (%/100 volts)	10
Minimum Dead Time (micro sec)	20
Gamma Sensitivity Co ⁶⁰ (cps/mR/hr)	60
Maximum Background Sheilded 50mmPb + 3mmAl (cpm)	30
Tube Capacitance (pf)	3
Weight (grams)	125


डी. गोपी / D. GOPI
 तकनीकी प्रबंधक / Technical Manager
 एन एस एस / आर आई डी / ई सी आई एल
 NSS / RID / ECIL
 हैदराबाद / HYDERABAD-500 080